

To: Luschek, Robert[Luschek.Robert@epa.gov]; Andrea Hamilton[andrea.hamilton@tceq.texas.gov]
Cc: Verhalen, Frances[verhalen.frances@epa.gov]; Belk, Ellen[Belk.Ellen@epa.gov]; Curran, Trisha[Curran.Trisha@epa.gov]
From: Bhattacharya, Dipanjana
Sent: Fri 8/14/2015 8:09:17 PM
Subject: RE: Data Loss Notification for several sites

I am during emergency response work on Gold King Mine release until next week.

Thanks,

Dipanjana Bhattacharya

Air Monitoring Section

US EPA Region 6

1445 Ross Ave., Suite 1200

6PD-Q

Dallas, TX 75202-2733

214-665-6753

From: Luschek, Robert
Sent: Friday, August 14, 2015 2:56 PM
To: Andrea Hamilton
Cc: Verhalen, Frances; Belk, Ellen; Curran, Trisha; Bhattacharya, Dipanjana
Subject: RE: Data Loss Notification for several sites

Hi Andrea,

Frances Verhalen has been selected as the Chief of the Air Quality Analysis Section. Could you please replace my name with hers. Thank you!

Rob Luschek

U.S. EPA Region 6 (6PD-A)

214-665-7148

luschek.robert@epa.gov

From: Andrea Hamilton [<mailto:andrea.hamilton@tceq.texas.gov>]

Sent: Friday, August 14, 2015 2:44 PM

To: Luschek, Robert

Cc: Bhattacharya, Dipanjana; Curran, Trisha; Belk, Ellen; Richard Chism; lindsey.jones@tceq.texas.gov; Weslee Copeland; shawn.mckinney@tceq.texas.gov; Patricia Delacruz; Holly.Landuyt_tceq.texas.gov; Sally Klein; Gary Goldman; Neal Penney; stephanie.ma_tceq.texas.gov; Erin McDonald; cindy.maresh_tceq.texas.gov; Melanie Hotchkiss; Felicia Powell

Subject: Data Loss Notification for several sites

Good afternoon,

I am writing to notify you of data loss (continuous) in excess of 120 hours for data for the SO2 sampler at Corpus Christi Tuloso, Ozone sampler at Mission, and NOy sampler at Houston Deer Park #2.

Continuous data loss:

AQS No.	Site Name	Monitor/ Parameter	Affected Range Dates		Reason for Data Loss
483550026	Corpus Christi Tuloso	SO2	8/2/2015	- 8/7/2015	Instrument malfunction. UV lamp circuit board replaced 8/7; subsequent calibration passed.
482150043	Mission	Ozone	8/6/2015	- 8/11/2015	Equipment malfunction. Sample line blockage cleared 8/11; subsequent calibration passed.
482011039	Houston	NOy, NO	7/28/2015	- 8/3/2015	Instrument malfunction. Instrument adjusted 8/3;

Deer Park #2

subsequent calibration
passed.

Please let us know if you require any additional details.

Andrea Hamilton

Monitoring Division

Data Quality Team

MC-165, PO Box 13087

Austin, TX 78711-3087

(512) 239-1672 Office

(512) 239-0867 Fax